

Product Name :
SAY Bolt Viscometer For Testing Lab for Bitumen Testing Lab

Product Code :
NBTE-M12



Description :

Say bolt Viscometer, Electrically Heated, ASTM D88, D244, AASHTO T72 for Bitumen Testing - for the empirical measurement of Say bolt Viscosity of petroleum products at specified temperatures between 70o F and 210o F. This is also used for determining the Saybolt Furol Viscosity of bituminous materials at temperatures of 250, 275, 300, 350, 400 and 450 F. It comprises one each of cylindrical Oil cup, Universal Tip, Furol Tip, Bath Fitted with immersion Heater mounted on a stand. Dimmer stat for temperature control, Stirrer with shield. Complete with insulated handle and thermometer support receiving flask, withdrawal tube, filter funnel, thermometer support for cup and circular spirit level. Suitable for operation on 230 V 50 Hz, Single Phase, A.C.

Optional Extras :

- Brass Oil Tube
- Receiving Flask, 60ml
- Universal and Furol Tip (Set of 2)
- Filter Funnel
- Thermometer Support
- Heating Coil
- Thermometer Type 18 F Range : 66o F to 80o F
- Thermometer Type 19 F Range: 94o F to 80o F
- Thermometer Type 19 F Range: 120o F to 148o F
- Thermometer Type 20F Range: 134o F to 148o F
- Thermometer Type 21 F Range : 174o F to 188o F
- Thermometer Type 22 F Range : 204o F to 218o F
- Thermometer Type 77F Range: 245o F to 265o F
- Thermometer Type 77F Range: 295o F to 315o F
- Thermometer Type 78 F Range: 345o F to 365o F

- Thermometer Type 80F Range : 395o F to 415o F
- Thermometer Type 81F Range: 445o F to 465o F

Naugra Export

Website: www.naugraexport.com, **Email:** sales@naugraexport.com

Address: 6148/6, Guru Nanak Marg, Ambala Cantt, Haryana, India, **Phone:** +91-0171-2643080, 2601773